

17 January 2019 – Update

DHB Service Accreditation

Following the identification that the Ministry of Health Band 1 basic hoop bed lever had a potential risk for head entrapment, it was replaced on the MOH List with a unit which includes an anti-entrapment bar and also a rope and cleat to secure the lever firmly against the mattress.


The supplier has also produced an installation and risk assessment document for inclusion with all Bed Levers supplied. Always consider a risk assessment when providing equipment.

Enable New Zealand consider that this equipment should no longer be available through the DHB Service Accreditation process.

The Service Accreditation option has therefore been removed from the List and from RTL.

The current MOH Band 1 Bed Lever information is below.

Band 1

Bed Lever		Band 1			
		SPECIFICATIONS Flat bottom with anti-entrapment bar, cord and cleat			
SUPPLIER		DIMENSIONS			
Tastech Services Ltd		290mm width of board			
BRAND		MAX USER WEIGHT			
Tastech Services Ltd		150kg			
SUPPLIER PRODUCT ID	JDE#	LOOP HEIGHT	BOARD LENGTH	PRICE BREAK	ORDERING CODE
BLF3-600	893380	550mm	600mm	\$50-\$250	B289
BLF3	893398	550mm	800mm	\$50-\$250	B290