

12 December 2018 – Update

Ministry of Health List Equipment Update

It has been identified that the current Band 1 basic hoop bed lever has a potential risk for head entrapment. Recommendations propose that this risk is eliminated with the addition of a vertical centre bar and the use of a mattress strap/cleat.


In conjunction with MOH, these recommendations have been accepted and a new model is being introduced effective immediately.

These should be available for selection through RTL shortly.

Due to changing the model, there may be a delay in providing this equipment as demand may exceed manufacturing capacity at times.

Details of the new models are as below:

Band 1

BEDROOM						
Bed Lever						Band 1
				SPECIFICATIONS Flat bottom with anti-entrapment bar, cord and cleat		
SUPPLIER				DIMENSIONS		
Tastech Services Ltd				290mm width of board		
BRAND				MAX USER WEIGHT		
Tastech Services Ltd				150kg		
SUPPLIER PRODUCT ID	JDE#	LOOP HEIGHT	BOARD LENGTH	PRICEBREAK	ORDERING CODE	DHB SERVICE ACCREDITATION
BLF3-600	893380	550mm	600mm	\$50-\$250	B289	Yes
BLF3	893398	550mm	800mm	\$50-\$250	B290	Yes